

Manufactured by:
Southern Medical (Pty) Ltd.
55 Regency Drive,
Route 21 Corporate Park,
Irene, Centurion, 0062, South Africa

P O Box 17198
Lyttleton, 0140
South Africa
Tel: +27 12 667 6243/4
Email: info@southmed.co.za



IMPORTANT: PLEASE READ

For detailed information on correct & safe device use, please consult the Surgical Manual.

DESCRIPTION:

The SPICCA (Southern Cervical Fusion Cages for Anterior Placement) devices are wedge-shaped, PEEK (ASTM F2026) anterior cervical interbody fusion cages. The devices are made available with Tantalum (ASTM F560) radiographic markers and some device variants are made available with a Titanium Coating (ASTM F1580). The cages have at least one central "bone pocket" that is intended to be filled with autograft or allogenic bone graft. The SPICCA, SPICCA-2 and SPICCA-F devices are intended to be used in combination with the AXIS anterior cervical plate system, while the SPICCA-S2 and SPICCA-SP devices are stand-alone devices that do not require supplemental fixation.

INDICATIONS FOR USE:

The SPICCA, SPICCA-2 & SPICCA-F Cervical Fusion Cages are cervical interbody fusion devices intended for spinal fusion procedures at one or two levels from the C2/C3 disc space to the C7/T1 disc space in skeletally mature patients with degenerative disc disease (defined as neck pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies) of the cervical spine. Implants are to be implanted via an open, anterior approach and packed with autograft or allogenic bone graft comprised of cancellous and/or corticocancellous bone graft. Patients must have undergone a regimen of at least six (6) weeks non-operative treatment prior to being treated with these devices. The implant is intended to be used in combination with an anterior cervical plating system.

The SPICCA-SP & SPICCA-S2 Stand-Alone Cervical Fusion Cages are stand-alone interbody fusion devices intended for spinal fusion procedures at one or two levels from the C2/C3 disc space to the C7/T1 disc space in skeletally mature patients with degenerative disc disease (defined as neck pain of discogenic origin with degeneration of the disc confirmed by history and radiographic studies) of the cervical spine. Implants are to be implanted via an open, anterior approach and packed with autograft or allogenic bone graft comprised of cancellous and/or corticocancellous bone graft. Patients must have undergone a regimen of at least six (6) weeks non-operative treatment prior to being treated with these devices. The implant is designed to accommodate two screws. Two screws should be used to ensure adequate fixation of the implant.

CONTRAINDICATIONS:

- Patients with known or probable intolerance to the materials used in the manufacture of this device
- Patients with infection, inflammation, fever, tumors, elevated white blood count, obesity, pregnancy, mental illness and other medical conditions which would prohibit a beneficial surgical outcome
- Patients resistant to following post-operative restrictions on movement especially in athletic and occupational activities.
- Use with components from other manufacturers.
- Grossly distorted anatomy caused by congenital abnormalities.
- Rapid joint disease, bone absorption or osteopenia. Osteoporosis is a relative contraindication since this condition may limit the degree of obtainable correction or stabilization, and/or the amount of mechanical fixation.
- Any case where the implant components selected for use would be too large or too small to achieve a successful result.
- Any patient having inadequate tissue coverage over the operative site or inadequate bone stock or quality.
- Any patient in which implant utilization would interfere with anatomical structures or expected physiological performance.
- Any case not described in the indications for use.
- Reuse or multiple use.
- Any other medical or surgical condition which would preclude the potential benefit of spinal implant surgery.

RISK & UNDESIRABLE SIDE-EFFECTS:

- Damage to the vertebral endplates and/or fracture of the vertebrae.
- Herniated nucleus pulposus, disc disruption, or degeneration at, above, or below the level of surgery.
- Soft tissue injury.
- Esophageal injury/perforation.
- Dural tears, persistent CSF leakage, or meningitis.
- Bone loss or decrease in bone density.
- Graft site complications (if autograft is used).
- Hemorrhage, hematoma, occlusion, seroma, edema, embolism, stroke, excessive bleeding, phlebitis, damage to blood vessels, or cardiovascular system compromise.
- Infection.
- Foreign body (allergic) reaction to implant.
- Post-operative change in spinal curvature, loss of correction, height and/or reduction.
- Loss of neurological function including paralysis, radiculopathy, and/or the development or continuation of pain, numbness, spasms, or sensory loss.
- Transient or permanent neurological deficits, reflex deficits, irritation, and/or muscle loss.
- Scar formation possibly causing neurological compromise around nerves and/or pain.
- Non-union (or pseudarthrosis), delayed union or mal-union.
- Disassembly, bending, and/or breakage of the implant.
- Implant loosening and/or migration.
- Subsidence of the device into the vertebral body(ies).
- Transient or persistent dysphagia (swallowing difficulties).
- Sore throat and/or hoarseness.
- Development of respiratory problems.
- Loss of spinal mobility or function.
- Wound necrosis or wound dehiscence.
- Inability to perform the activities of daily living.
- Change in mental status.
- Death.

Additional surgical intervention may be required to correct/prevent some of these possible adverse events.

WARNINGS & PRECAUTIONS: STERILITY, PACKAGING & RE-USE



The devices are sterilized via gamma irradiation and are provided **STERILE**. Do not re-sterilize the device. Re-sterilization could cause material degradation and could result in mechanical failure of the device, host rejection and/or post-operative infection.



Do not use implants if the packaging has been damaged or previously opened, or if the expiration date on the label has been exceeded.



The devices are **SINGLE-USE**. Do not re-use implants. An explanted implant must never be re-implanted. Re-use or re-implantation may result in cross-contamination or infection.



WARNINGS & PRECAUTIONS: CORRECT & SAFE USE

The device should only be implanted by qualified surgeons with knowledge of the correct and safe use of the device and associated instrumentation. Refer to the Surgical Manual provided by Southern Medical (Pty) Ltd. If uncertain, please contact a Southern Medical Representative.



The SPICCA, SPICCA-2 & SPICCA-F devices are intended to be used in combination with the AXIS anterior cervical plate system. Failure to use supplemental fixation may lead to mechanical failure of the device, non-union and/or cage migration/expulsion.

The SPICCA-SP & SPICCA-S2 devices are intended to be implanted with two screws provided by the manufacturer. Failure to use two screws may lead to mechanical failure of the device, non-union and/or cage migration/expulsion.

Autograft or allogenic bone graft should be used in combination with the device. Failure to use bone graft may result in non-union. In the absence of osseous fusion, mechanical failure of the implant can be expected as a result of everyday mechanical stresses.

STORAGE

There are no special storage instructions. Storage conditions must not lead to degradation or contamination of the packaging or its contents. Handle with care.

POST-OPERATIVE CARE INSTRUCTIONS

The surgeon/physician should provide appropriate postoperative care instructions to the patient and should ensure that the patient understands the potential consequences of non-compliance. To achieve the desired clinical outcome, the patient should be instructed to minimize physical activity as far as possible and avoid activities that require lifting, bending, twisting or excessive movement. Follow-up consultations should be arranged with the patient at appropriate intervals, as deemed necessary by the physician. The patient should only resume their normal activities once cleared to do so by their physician.



MRI SAFETY INFORMATION

The SPICCA devices have not been evaluated for safety in the MR environment. It has not been tested for heating or unwanted movement in the MR environment. The safety of the SPICCA devices in the MR environment is unknown. Performing an MR exam on a person who has this medical device may result in injury or device malfunction.

DEVICE REMOVAL

Surgical removal of the device is possible. A decision to remove the device must be made by a healthcare professional after taking into consideration the patient's general medical condition and the potential risk of a subsequent surgical procedure. Refer to the Surgical Manual for removal instructions.













DISPOSAL

Single-use devices that have been in contact with blood or bodily fluids/tissues should be disposed of in accordance with the hospital's procedure for disposal of hazardous and/or biological waste. Users must wear surgical gloves and take care to avoid sharp edges. Devices that cannot be used because the packaging has been damaged, but have not been in contact with blood or bodily fluids/tissues, should be returned to the manufacturer.

REPORTING

Any device-related adverse events must be reported to the manufacturer as soon as possible.

DESCRIPTIONS OF SYMBOLS USED IN PACKAGING

MANUFACTURER	USE BY
	
CONSULT INSTRUCTIONS FOR USE 	CATALOGUE NUMBER 
DO NOT RE-USE 	CAUTION 
DO NOT RE-STERILIZE 	LOT NUMBER 
DATE OF MANUFACTURE 	STERILIZED USING IRRADIATION 
DO NOT USE IF PACKAGING IS DAMAGED 	DOUBLE STERILE BARRIER 
CAUTION: FEDERAL LAW RESTRICTS THIS DEVICE TO SALE BY OR ON THE ORDER OF A PHYSICIAN 